



—Sales, Inc.—

225 Viscoloid Avenue  
Leominster, MA 01453

# CERTIFICATE OF QUALITY ASSURANCE

## **BIOLOG**

**PRODUCT CATALOG NO. 3102**

**LOT NO. 09624**

25 mL Modified Polystyrene Reagent Reservoir  
10 per package and 10 packages per case  
Sterile, RNase/DNase Free, Non-Pyrogenic

**MATERIAL:** 100 % VIRGIN FDA Grade Modified Polystyrene

**NON-PYROGENIC:** LESS THAN 0.03 EU/mL  
LAL GEL CLOT METHOD

**RNase/DNase FREE:** LOT CERTIFIED

GMP SALES, INC MAINTAINS ALL TEST RESULTS IN OUR FACILITY.  
ALL TESTS ARE PERFORMED BY AN INDEPENDENT CERTIFIED TESTING  
LABORATORY.

**STERILITY:** STERILIZED, Dose set to meet SAL  $10^{-6}$

QA APPROVED: VLADIMIR POWELL TITLE: QC MANAGER DATE: JUNE 2025  
EXPIRATION DATE: JUNE 2028