

# CERTIFICATE OF ANALYSIS

Product No.: 72401  
Name: Inoculating Fluid-A (IF-A)  
Lot No.: 25OCT091  
Expiration Date: 09 OCT 2027

---

## Quality Profile

---

<u>Challenge organisms</u>	<u>Source</u>	<u>Growth Result</u>
----------------------------	---------------	----------------------

No Growth Promotion Required

---

Final pH:  $6.9 \pm 0.4$  at 25 °C.  
Color: Light Green  
Unit: Tube  
Volume: 12 mL

Samples of this lot have met pH, appearance, and performance standards established by Biolog, Inc. Representative samples were incubated for 72 hours at room temperature, membrane filtration processed samples were incubated at 30 °C – 35 °C for 48 hours and found to be acceptable. Results were obtained at the time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

---

Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

---



Quality Assurance

30 OCT 2025

Date

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1 800 284 4949

CA72401 RevD  
Effective Date: 10/20/2023