

## CERTIFICATE OF ANALYSIS

**Catalog No.:** 72501  
**Product Name:** YT Inoculating Fluid  
**Lot No.:** 2603131  
**Expiration Date:** 13 MAR 2028

### Quality Profile

Parameter	Result
Appearance:	Passes test
Contamination check:	Passes test

Samples of this lot have met appearance and sterility standards established by Biolog, Inc. Representative samples were incubated at 26, 30 and 35 °C for 96 hours for contamination testing and found to be acceptable. Results were obtained at time of release. Certification does not exempt the user from his/her responsibility to assure satisfactory performance at the time of use.

Color: Clear, colorless liquid  
Unit: Tube  
Volume: 12.5 mL

### Intended Use

Not for human *in vitro* diagnostic use

For Research Use Only

DocuSigned by:  
Shawn Nette

**Quality Assurance**

3/27/2026

**Date**

**biolog**

21124 Cabot Blvd.  
Hayward, CA 94545

biolog.com  
+1800 284 4949

CA72501 RevC  
Effective Date: 10/20/2023