



—Sales, Inc.—

225 Viscoloid Avenue
Leominster, MA 01453

CERTIFICATE OF QUALITY ASSURANCE

BIOLOG

PRODUCT CATALOG NO. 3102

LOT NO. 17122

25 mL Modified Polystyrene Reagent Reservoir
10 per package and 10 packages per case
Sterile, RNase/DNase Free, Non-Pyrogenic

MATERIAL: 100 % VIRGIN FDA Grade Modified Polystyrene

NON-PYROGENIC: LESS THAN 0.03 EU/mL
LAL GEL CLOT METHOD

RNase/DNase FREE: LOT CERTIFIED

GMP SALES, INC MAINTAINS ALL TEST RESULTS IN OUR FACILITY.
ALL TESTS ARE PERFORMED BY AN INDEPENDENT CERTIFIED TESTING
LABORATORY.

STERILITY: STERILIZED, Dose set to meet SAL 10^{-6}

QA APPROVED: VLADIMIR POWELL TITLE: QC MANAGER DATE: NOVEMBER 2023
EXPIRATION DATE: NOVEMBER 2026